

# United Kingdom Accreditation Service

---

---

## ACCREDITATION CERTIFICATE



**CALIBRATION LABORATORY**  
**No. 0576**

**Kistler Instruments Ltd**

is accredited in accordance with the recognised International Standard ISO/IEC 17025:2005  
General Requirements for the competence of testing and calibration laboratories.

This accreditation demonstrates technical competence for a defined scope as detailed in and at the locations specified in the schedule to this certificate, and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communiqué dated January 2009).

The schedule to this certificate is an essential accreditation document and from time to time may be revised and reissued by the United Kingdom Accreditation Service. The most recent issue of the schedule of accreditation, which bears the same accreditation number as this certificate, is available from the UKAS website [www.ukas.com](http://www.ukas.com).

This accreditation is subject to continuing conformity with United Kingdom Accreditation Service requirements. The absence of a schedule on the UKAS website indicates that the accreditation is no longer in force.

A handwritten signature in blue ink, appearing to read 'M. Cooper', is written over a horizontal line.

*Accreditation Manager, United Kingdom Accreditation Service*

**Initial Accreditation date**  
**17 July 1998**

**This certificate issued on**  
**25 June 2012**


UKAS is appointed as the sole national accreditation body for the UK by The Accreditation Regulations 2009 (SI No 3155/2009) and operates under a Memorandum of Understanding (MoU) with the Department for Business, Innovation and Skills (BIS).

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>0576</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Kistler Instruments Ltd</b>	
	Issue No: 032      Issue date: 09 May 2018	
	13 Murrell Green Business Park London Road Hook Hampshire RG27 9GR	Contact: Mr Andy Harbour Tel: +44 (0)1256 741550 Fax: +44 (0)1256 741551 E-Mail: <a href="mailto:service.uk@kistler.com">service.uk@kistler.com</a> Website: <a href="http://www.kistler.com">www.kistler.com</a>
Calibration performed at the above address only		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
DC Charge	2 pC to 10 pC 10 pC to 100 pC 100 pC to 2000000 pC	0.27 % 0.15 % 0.14 %	For the calibration of charge amplifiers
DC Voltage	0 V to 10 V	0.060 % + 4.0 mV	For the calibration of charge amplifiers
END			